



Peri-Strips Dry
*with Veritas Collagen Matrix
Staple Line Reinforcement*

CONFIDENCE AT THE STAPLE LINE

Strong from the start. Strong over time.*

BARIATRIC PROCEDURES:

- ▶ Sleeve Gastrectomy
- ▶ Roux-en-Y
- ▶ Duodenal Switch (BPD/DS)

THORACIC PROCEDURES:

- ▶ Lung Volume Reduction
- ▶ Lobectomies
- ▶ Pneumonectomy



*Pre-clinical data results may not correlate to results in humans.

Baxter

Your staple line buttressing options are synthetic PGA or a **biological material**. Peri-strips Dry with Veritas stands apart as a biological material that is shown to decrease the risk of staple line bleeding and leaking over synthetic PGA.³

Applications & Procedures

PERI-STRIPS DRY reinforcement with VERITAS Collagen Matrix is derived from **bovine pericardium** and processed during manufacturing to produce a biocompatible, acellular, non-crosslinked collagen matrix that is immunologically inert!^{1,2*}

PERI-STRIPS DRY with Veritas Collagen Matrix is intended to reinforce the staple line during repair of soft tissue deficiencies and can be used during lung/bronchus resections, bariatric, gastric, and small bowel procedures.

Bariatric Procedures

Sleeve gastrectomy




Roux-en-Y




Thoracic Procedures

Lung volume reduction surgery



Lung resection




Strong Biological Reinforcement

PERI-STRIPS DRY with VERITAS Collagen Matrix is created from a proprietary tissue manufacturing technique designed to provide **high levels of biocompatibility** (based on histological features^{2*}) while maintaining the **intrinsic suppleness and strength** of the material!^{1*}

- Strength over time at the staple line vs. synthetic PGA material⁴
- Reduced risk of staple line bleeding and leaks^{5,6,7}
- Remodels in conjunction with the body's natural healing process⁸

Staple Line Reinforcement in Sleeve Gastrectomy Procedures



Gastric Staple Line Reinforcement - Clinical Meta-Analysis by Shikora et. al.⁹

The largest meta-analysis to date evaluating the outcomes for gastric staple line reinforcement was conducted through a systematic review of medical literature from 2000-2013.

There were **66,727** unique patients in the analysis. Data was extracted on **56,309** patients for leak and **41,864** patients for bleeding. These patients were included in 516 study arms, 301 for leak and 215 for bleed.


The three most commonly used SLR methods were compared

- 1 Oversewing reinforcement
- 2 Buttressing using PERI-STRIPS DRY / PERI-STRIPS DRY with VERITAS Collagen Matrix bovine pericardium
- 3 Buttressing using SEAMGUARD Bioabsorbable Staple Line Reinforcement glycolide copolymer PGA

16,967 articles were screened and **295** studies and abstracts were retained for data extraction.

Reinforcement type affects leak rate


Reduced leak rate in sleeve gastrectomy/gastric bypass patients



PSDV vs oversuture and no SLR: $p < 0.001$
PSDV vs SEAMGUARD: $p < 0.01$

Reinforcement type affects bleed rate

Reduced bleed rate in sleeve gastrectomy/gastric bypass patients



PSDV vs no SLR: $p < 0.001$
PSDV vs SEAMGUARD: $p < 0.01$
PSDV vs oversuture: $p \leq 0.05$
SEAMGUARD vs no SLR: $p < 0.01$

This large meta-analysis demonstrated that in a population predominantly comprised of morbidly obese bariatric patients, excluding high risk revision patients, PERI-STRIPS DRY / PERI-STRIPS DRY with VERITAS Collagen Matrix (bovine pericardium) outperformed no staple line reinforcement, suture oversewing and synthetic buttress, SEAMGUARD Bioabsorbable Suture Line Reinforcement.

Please see Indications and Important Risk Information on the back of this brochure.

The Risk of Leaks in Bariatric Surgery

The event of bleeding or leakage in gastrointestinal and bariatric surgeries can result in devastating complications for the patient and lead increased cost of care.⁹

Laparoscopic Sleeve Gastrectomy - Clinical Publication by D'Ugo et.al.⁵

Retrospective multicenter study of **1162** patients from 2002 to 2012 focused on the comparative use of different techniques to evaluate reducing leak and bleeding along the staple line during **laparoscopic sleeve gastrectomy (LSG)**.

Of the **1162** LSG patients, there were **305** males and **857** females with a mean age of 43.7 ± 9.4 years (range 17-67 yr) and a mean body mass index (BMI) of 48 ± 6.4 kg/m² (range 35-84 kg/m²)

Effects of reinforcing the staple line

- 1 The study indicated that staple line reinforcement with PERI-STRIPS DRY Staple Line Reinforcement with VERITAS Collagen Matrix significantly reduced the risk of postoperative leaks compared to all other groups in the study.
- 2 Postoperative bleeding was significantly lower with staple line reinforcement than without any buttressing along the staple line in LSG.

Among the total **1162** patients undergoing LSG, **973** utilized staple line reinforcement.

% Rate of Leak Episodes (n=1162)

0.3% PERI-STRIPS DRY w/ VERITAS

3.2% GORE

7.8% COVIDIEN

3.0% OVER-SEWING

4.8% NO SLR

% Rate of Bleeding Episodes (n=1162)

0% PERI-STRIPS DRY w/ VERITAS

1.6% GORE

1.3% COVIDIEN

1.4% OVER-SEWING

13.7% NO SLR


The authors of this study had no commercial associations that might be a conflict of interest.

Please see Indications and Important Risk Information on the back of this brochure.

Strong From the Start (Burst Strength)*

An experimental study evaluating staple line seam integrity at intraluminal pressures in the immediate postoperative period with a canine model demonstrated:

- Regardless of the operative techniques, the buttressed ileocolic anastomoses showed higher bursting pressures^{10*}
- Buttressed anastomoses never ruptured at the staple line^{10*}



Enhanced Reapproximation

Staple Reapproximation in Laboratory Testing.^{11 *}




SEM (scanning electron microscopy) shows the PERI-STRIPS DRY with VERITAS Collagen Matrix material to be a smooth, dense membrane which readily reapproximates around individual staples of a buttressed intestinal staple line. In contrast, the SEAMGUARD Bioabsorbable material is highly porous (metal staples, 25x mag).^{12*}

*Pre-clinical data results may not correlate to results in humans. Please see Indications and Important Risk Information on the back of this brochure.


Strong Over Time (Tensile Strength)*

Samples of PERI-STRIPS DRY with VERITAS Collagen Matrix and SEAMGUARD were placed in sterile saline and incubated at 37°C for 0, 7, 14, 21, or 28 days. **The study showed no significant decrease in tensile strength over time with PERI-STRIPS DRY with VERITAS.^{4*}**


- After 7 days in sterile saline at 37°C, both buttress materials showed no statistically significant change in tensile strength
- However, there was a significant decrease in tensile strength of the SEAMGUARD Bioabsorbable material between day 7 and day 14 and thereafter



Day 14 Tensile Strength^{4*}



Day 28 Tensile Strength^{4*}



Gore's SEAMGUARD is composed of a glycolide and trimethylene carbonate copolymer and claims to be absorbed into the body within six months of implantation.¹³ The comparative study above between PERI-STRIPS DRY with VERITAS and SEAMGUARD demonstrates that the ultimate tensile strength (UTS) of PERI-STRIPS DRY with VERITAS was always higher than SEAMGUARD over a 28 day time period. The ultimate tensile strength (UTS) is the maximum stress that a sample achieves throughout a tensile test and is generally the stress applied to the sample immediately before it fails.

*Pre-clinical data results may not correlate to results in humans. Please see Indications and Important Risk Information on the back of this brochure.

Remodeling vs. Resorption

Tissue Integration Through Remodeling

Unlike PERI-STRIPS DRY, which is a permanent, crosslinked collagen matrix, PERI-STRIPS DRY with VERITAS Collagen Matrix is a non-crosslinked matrix designed to remodel.^{8*}

Remodeling of PERI-STRIPS DRY with VERITAS Collagen Matrix entails two simultaneous processes that occur during the course of normal wound healing:^{8,14,15*}

- Degradation of the VERITAS Collagen Matrix scaffold
- Deposition of new tissue and blood vessels

Although degradation is an important aspect of the remodeling process, it is a spontaneous chemical process of decomposition that is not controlled by the body.¹⁴ **VERITAS Collagen Matrix exhibits favorable remodeling characteristics, such as scaffold degradation and new blood cell deposition, that occur during normal wound healing.**^{11*}

“Veritas acts as a scaffold that, with time, becomes infiltrated with host cells and supports angiogenesis.”
— B. Oray^{1*}

Replaced over time by tissue that is histologically indistinguishable from host tissue^{15*}



Time Zero

The acellular nature of the PERI-STRIPS DRY with VERITAS implant clearly distinguishes it from the surrounding tissue.


(Porcine model, 20x orig mag)



Three Days Post-Implant

Early signs of remodeling are evidenced by the infiltration of host cells into the PERI-STRIPS DRY with VERITAS acellular implant.

(Porcine model, 40x orig mag)



Six Weeks Post-Implant

The PERI-STRIPS DRY with VERITAS implant appears to be completely remodeled into histologically indistinguishable host tissue.

(Porcine model, 40x orig mag)

*Pre-clinical data results may not correlate to results in humans.
Please see Indications and Important Risk Information on the back of this brochure.

Preparation and Application

Preparation and Application

1

GEL BOTH STRIPS

Apply a continuous bead of PSD Gel to the center of both sides of the loading unit.



2

WEDGE LOADING UNIT

The plastic loading unit contains two strips of bovine pericardium secured on each side of a foam spacer. Gently push the PERI-STRIPS DRY with VERITAS Collagen Matrix loading unit into the jaws of the stapler and slide the foam insert out of loading unit.

Ethicon stapler



With the Covidien stapler, be careful to ensure that the end of the yellow guard is positioned above the assembly clip.

Covidien stapler



Note: PERI-STRIPS DRY with VERITAS can be used immediately or allowed to remain between the forks of the stapler loading unit until surgery. For optimal results, PERI-STRIPS DRY with VERITAS should be used within 60 minutes of its application to the stapler.

3

REMOVE PLASTIC SHEATH / HOLD CLOSED

Remove the plastic sheath by sliding it straight off in a continuous motion. Wipe off any excess gel from both sides of the foam insert. Leave the stapler locked on the foam insert until ready to use.

PERI-STRIPS DRY with VERITAS Collagen Matrix can be used immediately or allowed to remain between the jaws of the stapler loading unit until surgery. For optimal results, PERI-STRIPS DRY with VERITAS Collagen Matrix should be used within 60 minutes of its application to the stapler.



Please see Indications and Important Risk Information on the back of this brochure.

Primary Components

PERI-STRIPS DRY with VERITAS Collagen Matrix is composed of two primary components: the PERI-STRIPS DRY with VERITAS Collagen Matrix plastic loading unit and the PSD Gel. The plastic loading unit contains two strips of dehydrated bovine pericardium secured on each side of a foam spacer by the plastic loading unit. PSD Gel is a hydrogel which creates a temporary bond between the strips and the forks of the stapler and promotes rehydration of the strips.



PERI-STRIPS DRY with VERITAS Linear



PERI-STRIPS DRY with VERITAS Circular



PSD Gel Adhesive for PERI-STRIPS DRY

Defining Confidence at the Staple Line: PERI-STRIPS DRY with VERITAS Collagen Matrix



Ordering Information

	Endoscopic Linear Staplers	Size	Cartridge Color	PSDV Standard Order #	PSDV THIN Order #
ETHICON	ENDOPATH® ETS 45 Endoscopic Transecting	45	Gr Wh Bl Gr	PSD4506ETSV	PSD4506ETSVBIO
	ENDOPATH® ETS Flex 45	45	Gr Wh Bl Gr	PSD4506ETSV	PSD4506ETSVBIO
	ENDOPATH® EZ 45	45	Bl Gr	PSD4506EV	
	ENDOPATH® ETS Flex 60	60	Wh Bl Gr	PSD6006ETSV	
	Echelon 45 ENDOPATH® Stapler	45	Gr Wh Bl Gl Gr	PSD4506ECHV	PSD4506ECHVBIO
	Echelon 60 ENDOPATH® Stapler	60	Wh Bl Gl Gr Bl	PSD6006ECHV	PSD6006ECHVBIO
	Echelon FLEX 45 ENDOPATH®	45	Gr Wh Bl Gl Gr	PSD4506ECHV	PSD4506ECHVBIO
	Echelon FLEX 60 ENDOPATH®	60	Wh Bl Gl Gr Bl	PSD6006ECHV	PSD6006ECHVBIO
COVIDIEN	Endo GIA™ Universal Straight/Rotulator 30	30	Gr Wh Bl	PSD3006UV	
	MultiFire Endo GIA™ 30	30	Gr Wh Bl	PSD3006UV	
	Endo GIA™ Reloads with Tri-Staple™ Technology	30	Gr Ta Pu	PSD3006UV	
	Endo GIA™ Universal Straight 45	45	Gr Wh Bl Gr	PSD4506UV	PSD4506UVBIO
	Endo GIA™ Universal Rotulator 45	45	Wh Bl Gr	PSD4506UV	PSD4506UVBIO
	Endo GIA™ Reloads with Tri-Staple™ Technology	45	Bl Pu	PSD4506UV	PSD4506UVBIO
	Endo GIA™ Universal Straight/ Rotulator 60	60	Wh Bl Gr	PSD6006UV	PSD6006UVBIO
	Endo GIA™ Reloads with Tri-Staple™ Technology	60	Bl Pu	PSD6006UV	PSD6006UVBIO
INTUITIVE	EndoWrist Stapler 30 for da Vinci Xi	30	Gr Wh Bl Gr	PSD3006UV	
	EndoWrist Stapler 45 for da Vinci Xi	45	Wh Bl Gr	PSD4506ETSV	PSD4506ETSVBIO
	EndoWrist Stapler 45 for the da Vinci Si	45	Bl Gr	PSD4506ETSV	PSD4506ETSVBIO

	Open Linear Staplers	Size	Cartridge Color	PSDV Standard Order #	Cartridge Color Key
ETHICON	PROXIMATE® Linear Cutter 55	55	Wh Bl Gr	PSD5506EV	Wh White
	USSC ILA 52	52		PSD5506EV	Bl Blue
	PROXIMATE® Linear Cutter 75	75	Bl Ye Gr	PSD7506EV	Gl Gold
	USSC ILA 75	75		PSD7506EV	Gr Green
COVIDIEN	PROXIMATE® Linear Cutter 100	100	Bl Gr	PSD10006EV	Bl Black
	GIA™ DST 60	60	Wh Bl Gr	PSD6006UDSTV	Ye Yellow
	GIA™ DST 80	80	Bl Gr	PSD8006UDSTV	Pu Purple
	GIA 80	80		PSD8006UV	Gr Gray
	GIA 90	90		PSD9006UV	Ta Tan
	PI 90	90		PSD9006PIV	
	GIA™ DST 100	100	Bl Gr	PSD10006EV	
E	Circular Staplers	Size	Cartridge Color	PSDV Standard #	
	PROXIMATE ILS 25mm, ENDOPATH® ILS 25mm	25		PSD25V	
	DST EEA25mm, ORVIL Device 25mm	25		PSD25V	
C	Premium Plus CEEA™ 25	25	Wh	PSD25V	

TO ORDER CALL: 1-888-229-0001 Option #3



Peri-Strips Dry

with Veritas Collagen Matrix
Staple Line Reinforcement

PERI-STRIPS DRY Indications for Use:

PERI-STRIPS DRY with VERITAS (PSDV) is intended for use as a prosthesis for the surgical repair of soft tissue deficiencies using surgical staplers when staple line reinforcement is needed.

PSDV can be used for reinforcement of staple lines during lung and bronchus resections and during bariatric surgical procedures.

PSDV can be used for reinforcement of staple lines during gastric, bariatric, and small bowel procedures.

Contraindications:

The use of PSDV is contraindicated in patients with known sensitivity to bovine or acrylic material.

Adverse Reactions:

As with any surgical procedure, adverse reactions are possible and include but are not limited to: infection, rejection, erosion, and allergic reaction.

PSD Gel Indications for Use:

PSD Gel is intended to be used only with Peri-Strips Dry (PSD) as a temporary bonding agent during applications indicated in the PSD Instructions for Use.

PSD Gel Contraindications:

PSD Gel is not designed, sold, or intended for use except as indicated. Use only as a temporary bonding agent with Peri-Strips Dry.

Rx Only. For safe and proper use of this device refer to the complete Instructions for Use.

References:

1. Oray BN, Kelly S, Konobeck T, Lambert A, Mooradian DL. Novel propylene oxide-treated bovine pericardium as soft tissue repair material and potential scaffold for tissue engineering. *Surgical Technology International XVIII*. 2. Data on file. Baxter Healthcare Inc. 3. Spector D, Perry Z, Konobeck T, Mooradian D, Shikora S. Comparison of hemostatic properties between collagen and synthetic buttress materials used in staple line reinforcement in a swine splenic hemorrhage model. *Surg Endosc*. 2011;25(11):148-1152. 4. Data on file. SI-0107. Synovis Surgical Innovations. 5. D'Ugo S, Gentileschi P, Benavoli D, et al. Comparative use of different techniques for leak and bleeding prevention during laparoscopic sleeve gastrectomy: a multicenter study. *Surg Obes Relat Dis*. 2013 Nov 12. [Epub ahead of print] 6. Stamou KM, Menenakos E, Dardamanis D, et al. Prospective comparative study of the efficacy of staple line reinforcement in laparoscopic sleeve gastrectomy. *Surg Endosc*. 2011;25(11):3526-3530. 7. Shikora SA, Kim JJ, Tarnoff ME. Comparison of permanent and nonpermanent staple line buttressing materials for linear gastric staple lines during laparoscopic Roux-en-Y gastric bypass. *Surg Obes Relat Dis*. 2008;4(6):729-734. 8. Deeken C, Melman L, Jenkins E, Greco S, Frisella M, Matthews B. Histologic and biomechanical evaluation of crosslinked and non-crosslinked biologic meshes in a porcine model of ventral incisional hernia repair. *J Am Coll Surg*. 2011;212(5):880-888. 9. Shikora SA, Mahoney CB. Clinical benefit of gastric staple line reinforcement (SLR) in gastrointestinal surgery: a meta-analysis. *Obes Surg*. 2015. Published online; doi 10.1007/s11695-015-1703-x 10. Gaertner WB, Hagerman GF, Potter MJ, Karulf RE. Experimental evaluation of a bovine pericardium-derived collagen matrix buttress in ileocolic and colon anastomoses. *J Biomed Mater Res B Appl Biomater*. 2010;92(1):48-54. 11. Data on file. Synovis Surgical Innovations 12. Data on file. AS-0012R. Synovis Surgical Innovations. 13. Gore Medical. Instructions For Use For: GORE SEAMGUARD Bioabsorbable Staple Line Reinforcement. Mar. 2008. 14. Melman L, Jenkins E, Hamilton NA, et al. Early biocompatibility of crosslinked and non-crosslinked biologic meshes in a porcine model of ventral hernia repair. *Hernia*. 2011;15(2):157-164. 15. Hsu A, Mustoe TA. The principles of wound healing. Elsevier website. Available at: <http://www.elsevieradvantage.com/samplechapters/9780323034708/9780323034708.pdf>. Accessed May 22, 2015. 16. Data on file. AS-0004R. Synovis Surgical Innovations.

Baxter, Peri-Strips, Peri-Strips Dry and Veritas are registered trademarks of Baxter International Inc., or its subsidiaries.

Covidien is a registered trademark of Covidien AG.

Ethicon is a registered trademark of Johnson & Johnson Corporation.

All other products or trademarks appearing herein are the property of their respective owners.

USMP/390/15-0013(3) 12/2017

Baxter

Baxter International Inc.
One Baxter Parkway
Deerfield, Illinois 60015

www.baxter.com